

Table 1: Common implants and their MRI safety

Implant	MR Safety
Cardiac Pacemakers and Pacing wires	Contraindicated for MRI
Cardiac Defibrillators	Contraindicated for MRI
Intracranial Aneurysm Clips	Certain types are safe for MRI. Information about the make and model is needed.
Cochlear Implants	Certain types are safe for MRI. Information about the make and model is needed.
Intravascular Coils and Stents	Safe for MRI six weeks post-surgery
Heart Valves and Annuloplasty Rings	Certain types are safe for MRI. Information about the make and model is needed.
Endoscopic Clips	Contraindicated for MRI
Implanted Insulin Pumps	Contraindicated for MRI
Intrauterine Devices (IUD)	Most are safe for MRI (e.g. Mirena). Information about the make and model is needed.
Metallic Orbital Foreign Bodies	Contraindicated for MRI
Neuro-stimulators	Contraindicated for MRI
Ocular Implants	Certain types are safe for MRI. Information about the make and model is needed.
Orthopedic Implants	Safe for MRI six weeks post-surgery
Otologic (Ear) Implants	Certain types are safe for MRI. Information about the make and model is needed.
Penile Implants	Certain types are safe for MRI. Information about the make and model is needed.

You may find more useful information on the following website:

<http://www.mrisafety.com/>

You're always welcome to contact us if you have any question about the MR safety of your implant.